



## AMMONIA NITROGEN Kit

OCTA-SLIDE METHOD, 1-8 ppm

CODE 4795-01

QUANTITY	CONTENTS	CODE
30 mL	Ammonia Nitrogen Reagent #1	4797WT-G
30 mL	*Ammonia Nitrogen Reagent #2	*4798WT-G
2	Test Tubes, 2.5-10 mL, plastic, w/caps	0106
1	Octa-Slide 2 Viewer	1101
1	Ammonia Nitrogen Octa-Slide 2 Bar, 1.0-8.0 ppm	4796-01


\*WARNING: Reagents marked with an \* are considered to be potential health hazards. To view or print a Safety Data Sheet (SDS) for these reagents go to [www.lamotte.com](http://www.lamotte.com). Search for the four digit reagent code number listed on the reagent label, in the contents list or in the test procedures. Omit any letter that follows or precedes the four digit code number. For example, if the code is 4450WT-H, search 4450. To obtain a printed copy, contact LaMotte by email, phone or fax.

Emergency information for all LaMotte reagents is available from Chem-Tel (US, 1-800-255-3924) (International, call collect, 813-248-0585).

To order individual reagents or test kit components, use the specified code number.









WARNING! This set contains chemicals that may be harmful if misused. Read cautions on individual containers carefully. Not to be used by children except under adult supervision.

### USE OF THE OCTA-SLIDE 2 VIEWER



The Octa-Slide 2 Viewer should be held so non-direct light enters through the back of the Viewer. Insert the Octa-Slide 2 Bar into the Viewer. Insert the reacted sample into the top of the Viewer. Match the color of the reaction to the color standards.

# PROCEDURE



## LaMOTTE COMPANY

Helping People Solve Analytical Challenges

PO Box 329 · Chestertown · Maryland · 21620 · USA  
800-344-3100 · 410-778-3100 [Outside U.S.A.] · Fax 410-778-6394

Visit us on the web at [www.lamotte.com](http://www.lamotte.com)